


## Rapid RT-PCR Cycler Protocol



Step	Time	Temperature	No. of cycles
Reverse Transcription	8 min	48°C	1 cycle
Polymerase activation	2 min	96°C	1 cycle
Denaturation	2 sec	95°C	42 cycle
Annealing + Extension	10 sec + reading	60°C	

If you have further questions please contact our Service Team:

Phone.: +49 6404 – 925125

Email: [complaint@bag-diagnostics.com](mailto:complaint@bag-diagnostics.com)